



CASE STUDY

ADVANCED DIAGNOSTIC AIDS AND ITS FUTURE IN ORAL PATHOLOGY

*Dr. Hema Keswani, Dr. Yogesh T Lakkasetty, Dr. Hemavathy S. and Dr. Sarvesh Dave

Department of Oral and Maxillofacial Pathology, Sri Rajiv Gandhi College of Dental Sciences and Hospital,
Bangalore-560032, Karnataka, India

ARTICLE INFO

Article History:

Received 17th October, 2016
Received in revised form
25th November, 2016
Accepted 10th December, 2016
Published online 31st January, 2017

Key words:

Advancement,
Diagnosis, Oral cancer.

ABSTRACT

The pathology has extended its wings in the past few decades and has contributed greatly in understanding the pathogenesis of genetic disorders and in diagnosis of several undifferentiated malignant neoplasms. Molecular techniques are being used in clinical field. There are various techniques which are introduced in the field of pathology like brush cytology, velscope, confocal microscopy, tumor markers, microarray etc. New emerging technologies, including robotics, humanoid technology, lab-on-chip devices, nanodevices and patient 'smart' implants, will in the future offer unique opportunities for laboratories to develop. In present article few of the various techniques have been discussed which when used optimally by the pathologist have improved the quality of life of patients and newer future diagnostic techniques which could be boon in the field of pathology.

Copyright ©2017, Hema Keswani et al. This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

Citation: Dr. Hema Keswani, Dr. Yogesh T Lakkasetty, Dr. Hemavathy S. and Dr. Sarvesh Dave, 2017. "Advanced diagnostic aids and its future in oral pathology", *International Journal of Current Research*, 9, (01), 45704-45707.

INTRODUCTION

The first microscopic description of human tumors were found in a text by Sir Edward Home in 1830, followed by Johannes Muller in 1838. The term "biopsy" was coined by Ernest Besnier in 1879. There was restricted application of biopsy until the mid-20th century. (Medhini Singaraju *et al.*, 2013; Underwood, 1987) Wax embedding of specimens did not come into general use till the 1800s and histotechnology was still not in function. Frozen sections were being introduced in American medical centres in twentieth century. (Medhini Singaraju *et al.*, 2013; Underwood, 1987) According to Pearse a French botanist Francois-Vincent Raspail was the first to demonstrate the chemical reaction with the microscopic observation of tissues and cells. (Medhini Singaraju *et al.*, 2013; Pearse, 1980)

Vital Tissue Staining – Toluidine Blue Staining & Lugol's Iodine

Early malignant lesions show affinity for toluidine blue dye. Lugol's iodine and toluidine blue have been used together in the detection of early carcinomas and other oral lesions. Toluidine blue is a (acidophilic) metachromatic dye which selectively stains acidic tissue components, thus staining DNA and RNA. It stains mitochondrial DNA, cells with greater than

normal DNA content or altered DNA seen in dysplastic and malignant cells. Lugol's solution is used for delineation of the malignant change which produces a brown black stain when the iodine reacts with the glycogen content. The use of toluidine blue and Lugol's iodine serves as a useful adjunct in the diagnosis of patients who are at risk and for selecting the site for biopsy with wide field cancers prior to treatment. (Underwood, 1987; Masthan *et al.*, 2012) Toluidine blue has also been reported as an aid in selecting biopsy sites and in delineating the margins of lesions (Chaudhary *et al.*, 2014). According to recent studies TB stained lesions reported a link between carcinoma and loss of heterozygosity at 3p and 17p, while dysplasia resulted in loss of heterozygosity at 9p. The presence of loss of heterozygosity has also been reported in high frequency of TB-stained lesions without or with low grade dysplasia.

Chemoluminescence – Vizilite

This device has been approved for use in the United States by the Food and Drug Administration (November 2001). It has been investigated that the use of vizilite is beneficial as compared to Conventional visual examination. This is a hand-held, single-use, disposable chemiluminescent light stick that emits light at 430, 540 and 580 nm wavelengths. Normal epithelium will absorb light and appear dark whereas hyper keratinized or dysplastic lesions appear white. (Chaudhary *et al.*, 2014; Joel *et al.*, 2002; Diana V Messadi, 2013) It is recommended to use dilute acetic acid rinse and observe under a chemiluminescent light such as vizilite for improved early

*Corresponding author: Dr. Hema Keswani,

Department of Oral and Maxillofacial Pathology, Sri Rajiv Gandhi College of Dental Sciences and Hospital, Bangalore-560032, Karnataka, India.

detection of oral cancer (Masthan *et al.*, 2012). A recent study of high risk patients showed that the majority of lesions with a histological diagnosis of dysplasia or carcinoma in situ were detected and mapped using vizilite with TB (Joel *et al.*, 2002).

Brush cytology (Oral CDX)

Oral cdx is based on the concept of exfoliative cytology in the assessment of dysplastic changes in various suspected lesions especially in oral cancer. The use of brush cytology is inexpensive, simple, noninvasive and also risk-free technique. The oral epithelial cells can be obtained by the use of a cytobrush (thenon lacerational device), samples are fixed onto a glass slide, stained with a modified Papanicolaou test and analyzed microscopically via a computer-based imaging system. With brush cytology, sensitivity for detecting oral epithelial dysplasia but is associated with controversies and false negative results. (Underwood, 1987; Chaudhary *et al.*, 2014; Masthan *et al.*, 2012)

Velscope (narrow emission tissue fluorescence)

It has been used in screening and diagnosis of precancer and early cancer of lung, uterine cervix skin and oral cavity. (Diana V Messadi, 2013) The concept of tissue autofluorescence in diagnosis of dysplastic lesions of the oral cavity is based on changes in the structure and metabolism of the epithelium and the subepithelial stroma when interacting with light. Specifically, loss of autofluorescence in dysplastic and cancerous tissue is believed to reflect a complex mixture of alterations to intrinsic tissue fluorophore distribution, due to tissue remodeling such as the breakdown of the collagen matrix and elastin composition. Normal mucosa emits pale green autofluorescence. (Underwood, 1987; Diana V Messadi, 2013; Masthan, 2012). It act as a adjunct in improving distinction benign and malignant changes and identifying the malignant lesion that are not visible to naked eye under white light. (Diana V Messadi, 2013)

Confocal microscopy

This is an imaging technique for various researches in cell biology with an advantage of optical sectioning and high resolution imaging. This technique is useful in identifying the characteristic features such as nuclear irregularity which is used to differentiate OSCC from normal oral mucosa. However, further optimization of the instrument is still required. (Masthan, 2012)

Optical Coherence Tomography (OCT)

It was first reported by Fujimoto *et al.* (1991), it has numerous applications gastroenterology, ophthalmology, dermatology, and dentistry. OCT is a non-invasive, non-radiative optical diagnostic tool based on interferometers. (Underwood, 1987; Gaikwad *et al.*, 2013) There are few indicators involved in diagnosis of oral cancer, like the EP layer thickness and the standard deviation (SD) of OCT signal intensity. In an abnormal oral EP containing dysplastic cells, the cell size, shape, nucleus size, and arrangement is more randomly distributed as compared to healthy oral epithelium (EP). In this scenario, light scattering becomes stronger and its spatial distribution becomes more strongly fluctuated (Gaikwad *et al.*, 2013) The advantage of OCT is cross-sectional images of

normal or abnormal tissues can be obtained without biopsy and there is no exposure of the patient to ionizing radiation.

Tumor Markers & Bio Markers

Tumor markers may be present in blood circulation, body cavity fluids, cell membranes and cell cytoplasm. When released by cancer cells or produced by the host in response to cancerous substances. (Masthan *et al.*, 2012) The most reliable biomarker in OSCC development include the TSG p53 protein expression, chromosomal polysomy (DNA ploidy), and changes (termed loss of heterozygosity; LOH) in chromosomes 3p or 9p (probably due to changes in the TSG p16). Tumour Suppressor Genes, Oncogenes, cell proliferation markers, angiogenic markers and cell adhesion molecules are useful in prediction for the prognosis of patients with OSCC. (Masthan *et al.*, 2012) Molecular analysis of exfoliated cells also has shown the same changes as are present in tumor biopsy specimens. (Chaudhary *et al.*, 2014)

PCR-Based diagnostic aids

It can be used to detect mutations in cancer-associated oncogenes (e.g., K-ras, Nras), tumor suppressor genes (e.g., p53, p16) etc. and aids as an important detection tool. With PCR technique the range and sensitivity of diagnostic procedure has been increased. But still has a major drawback, of contamination and amplification artefacts may give rise to difficulties in the interpretation of the desired results. (Masthan *et al.*, 2012)

IDENTAFI 3000

This technology is a combination of anatomical imaging with fluorescence, fiber optics and confocal microscopy to map and delineate precisely the lesion in the area being screened. It is small in size and easy accessibility to all tissues in the oral cavity. The mechanism is similar to veloscope and also detect changes in angiogenesis with green amber light illumination. (Joel B. Epstein *et al.*, 2002)

Microarray

The pattern of gene expression vary in normal tissue and its malignant counterpart, which can be assessed by Tissue microarray technique. (Nigam *et al.*, 2014) RNA from malignant lesion and from a control tissue are extracted, and cDNA are prepared by reverse transcription. These reverse transcription reactions incorporate different fluorescent dyes so that cDNA from the tumor and from the control tissue can be distinguished by their fluorescence emission spectra. These cDNAs are mixed and then hybridized to the microarray. A relative increase in expression of a particular gene in a tumor sample leads to an increase in binding of labeled tumor-derived cDNA to the spot in the array that is complementary to the gene of interest. DNA microarrays are being used to detect single nucleotide polymorphisms (snps) of our genome (Hap Map Project), aberrations in methylation patterns, alterations in gene copy number, alternative RNA splicing, and pathogen detection. (Nigam *et al.*, 2014; Trevino *et al.*, 2007)

Future of diagnostic techniques

LAB ON CHIP- lab-on-a-chip or micro-total-analysis systems (TAS) also known as Microfluid technology. It is the

adaptation, miniaturization, integration and automation of analytical laboratory procedures into a single device or “chip”. Microfluidics is often regarded as the chemistry or biotechnology equivalent of the silicon integrated silicon chip that has revolutionized electronics, computers and communications. Microfluidics are by definition suited for handling living cells (whose typical diameter is a few micrometers) in a three-dimensional, biologically relevant environment. This microfluidic chip accepts saliva sample, can be operated by minimally trained personnel, and can provide a diagnostic answer in an automated and timely fashion. The detection of oral pre-cancer (dysplastic) and cancer cells within the chip will take advantage of membrane-associated cell proteins that are singularly expressed on cell cancer cells. (Govindarajan *et al.*, 2012; Mielczarek *et al.*, 2016)

Nuclear magnetic resonance microscopy

This will allow cellular pathologist to look is pattern of cell in tissue and facilitates examination of cells for the presence or absence or mutation of genes that control growth and function and will facilitate examination of specific marker of disease. This will allow non invasive 3D visualization of single in cell in living tissue. New innovation will be cellular metabolo imaging, cytonmr. (Kersaudy-Kerhoas and Sollier, 2013; Komoroski *et al.*, 1991)

Clinical microbiology

Microbiology and infectious disease will rely on lab on chip devices, including automated DNA, RNA and protein/peptide extraction chips coupled to organism identification chips and sequencing chips giving real time analysis of patient specimens. (Bock, 1989; Green, 2005)

Cyogenetics

The use of interphase cytogenetic fine gene locus mapping and locus specific sequencing of novel disease loci in patient with specific monitoring of locus specific changes following the treatment. (Saiki *et al.*, 1985; Chial, 2008)

Diagnostic molecular pathology

Diagnostic genomics

- Transcriptomes
- Polysonics
- Proteomics
- Peptidomics
- Pharmacogenomics
- Metabologenomics

Circulating tumor cells- (CTC assay)

Assay counts rare events – epithelial tumor cells in peripheral blood stream and compare to established frequency profiles. May predict treatment response more quickly than usual clinical practice with radiographic imaging (2-3 days vs 2-3 months) allowing rapid therapy modification. FDA approved for patients with metastatic breast cancer, tool for predicting progression and overall survival, monitoring disease progression. (Trask, 2002; Krebs *et al.*, 2010)

Dried blood spot

The test uses a dried blood spot specimen. It works by converting protein to peptides and then using a mass spectrometer to select and accurately measure diagnostic metabolites and /or peptides. It requires only a tiny blood sample. Genetic testing and molecular diagnostic assay can be performed from DBS specimen.^(21,22)

Virtual colonoscopy

It uses CT technology as an alternative to optical screening colonoscopy. VC digitally reconstructs the CT image into 2D and 3D pictures of colonic luminal surfaces.

Conclusion

In the future hopefully, pathologist will not be merely “tissue sampler”. Research will be focused on cancer prognostic markers and cancer cures via targeted therapies. Advances in molecular taxonomy will be an important adjunct to histologic diagnoses. Development of tissue banks, construction of microarrays, molecular research will be the framework in future of research pathology. Advances in genetics, information technology and digital imaging are already transforming histopathology and many other pathology specialities. Much interpretive reporting will be done from flat screens rather than through microscope eyepieces. Molecular diagnostic, genomics and proteomics will have the high impact because they will continue to redefine disease at the molecular level. The further development of nanotechnologies will also drive change and expedite diagnoses. The other changes will be increased automation; more things can be done with greater consistency, less time and less cost. In the developed world, pathology will still use many of the current techniques but these will be enhanced due to development predominantly in molecular pathology, information technology and imaging techniques and by commercialisation of new techniques and better instrumentation.

REFERENCES

- Alunni-Fabroni M., Sandri M.T. 2010. Circulating tumour cells in clinical practice: methods of detection and possible characterization. *Methods*, 50: 289–297
- Bock JL. 1989. Nuclear magnetic resonance in the clinical laboratory. *Am J Clin Pathol.*, Apr;91(4 Suppl 1):S19-26.
- Chaudhary R, Shah A, Shah DM, Singh S, Thakkar P, Goyal S. 2014. Advanced Diagnostic Aids in Detection of Oral Cancer. *Int J Dent Med Res.*, 1(3):139-143.
- Chial, H. 2008. Cytogenetic methods and disease: Flow cytometry, CGH, and FISH. *Nature Education*, 1(1):76
- Diana V Messadi, 2013. Diagnostic aids for detection of oral precancerous conditions, *International Journal of Oral Science*, 5, 59–65
- Gaikwad. P, Hiremath S K S, Singh Shhreddha, 2013. Advancement in Diagnostic Aids for Premalignant lesions- A Review. *Journal of Dental Sciences and Oral rehabilitation*, Jan-March issue 1;11-15
- Govindarajan, R., Duraiyan, J., Kaliyappan, K., & Palanisamy, M. 2012. Microarray and its applications. *Journal of Pharmacy & Bioallied Sciences*, 4(Suppl 2), S310–S312.
- Green, D.M. 2005. Improving health care and laboratory medicine: the past, present, and future of molecular diagnostics. *Proc. Bayl. Univ. Med. Cent.*, 18: 125–129.

- Joel B. Epstein, Lewei Zhang, 2002. Miriam Rosin Advances in the Diagnosis of Oral Premalignant and Malignant Lesions *J Can Dent Assoc.*, 68(10):617-21
- Kersaudy-Kerhoas M. and E. Sollier, Micro-scale blood plasma separation: from acoustophoresis to egg-beaters 2013. *Lab Chip*, 13, 3323–3346.
- KMK Masthan, N Aravindh Babu, Kailash Chandra Dash, M Elumalai, Advanced Diagnostic Aids in Oral Cancer Asian Pacific J Cancer Prev, 2012, 13(8); 3573-3576
- Komoroski RA, Pappas A, Hough A. 1991. Nuclear magnetic resonance in pathology: I. Principles and general aspects. *Hum Pathol.*, Nov; 22(11):1077-84.
- Krebs, M. G., Hou, J.-M., Ward, T. H., Blackhall, F. H. & Dive, C. 2010. Circulating tumour cells: their utility in cancer management and predicting outcomes. *Therapeutic Advances in Medical Oncology*, 2(6), 351–365.
- Lakshmy, R. 2008. Analysis of the Use of Dried Blood Spot Measurements in Disease Screening. *Journal of Diabetes Science and Technology*, 2(2), 242–243.
- Liang, Z. and Richards, R. 2010. Virtual colonoscopy vs optical colonoscopy. *Expert Opinion on Medical Diagnostics*, 4(2), 159–169.
- Medhini Singaraju, Sasidhar Singaraju, Kishore Sonawane, Sandeep Jain, Karthik B Gowda, 2013. Role of Advanced Diagnostic Aids in Oral Pathology, *Indian J Dent Adv.*, 5(1): 1125-1129
- Mielczarek, W. S., ab E. A. Obaje, b T. T. Bachmann and M. Kersaudy-Kerhoas : *Lab Chip*, 2016, 16, 3441
- Nigam P, Prasad K, Tak J, Gupta V, Sinha A, Bali R, Grewal P. 2014. Advanced Diagnostic Aids in Early Detection of Oral Cancer. *J Adv Med Dent Scie Res.*, 2(3):39-43
- Pearse AGE. 1980. Historical introduction. In *Histochemistry, theoretical and applied*, vol 1, Edinburgh, Churchill Livingstone.
- Saiki RK, Scharf S, Faloona F, Mullis KB, Horn GT, Erlich HA, Arnheim N. 1985. Enzymatic amplification of beta-globin genomic sequences and restriction site analysis for diagnosis of sickle cell anemia. *Science*, 230:1350–1354
- Solomon et al. 2002. Dried blood spots (DBS): a valuable tool for HIV surveillance in developing/tropical countries. *Int J STD AIDS.*, Jan; 13(1):25-8.
- Trask, B. J. 2002. Human cytogenetics: 46 chromosomes, 46 years and counting. *Nature Reviews Genetics*, 3, 769–778.
- Trevino, V., Falciani, F., Barrera-Saldana, H. 2007. A. DNA Microarrays: a Powerful Genomic Tool for Biomedical and Clinical Research. *Mol Med. Sep-Oct; 13(9-10):527-41.*
- Underwood JCE, editor: *Introduction to biopsy interpretation and surgical pathology*, ed 2, London, 1987, Springer-Verlag.
